

Development, Implementation and Audit of Locally Agreed Standards for Permanent Pacemaker Follow-Up

Presenting Author: E. Beaney, Highly Specialised Cardiac Physiologist
The James Cook University Hospital, Middlesbrough, UK


Authors: E. Beaney, Z. Patterson, M. Chapman, A.J. Turley, M.G.D. Bates, A.R. Thornley, S. James, N.J. Linker

Outline

- Why is the quality assurance (QA) process needed?
- QA standards
- Data collection
- Results
- Conclusions

Why is QA needed?

- QA identifies variation in clinical practice
- No comprehensive standards
 - HRS/EHRA¹
 - BHRS²
- BHRS recommend a formal QA process
- Suboptimal programming common³



- 1) Wilkoff et al. Europace 2008
- 2) BHRS Council, 2015
- 3) Ziacchi et al. Europace 2017

QA Standards

- **Patient symptoms documented ?**

- Symptoms potentially ppm-related thoroughly investigated?
- If heart failure symptoms reported, has appropriate referral been made or investigation been performed?
- Appropriate advice given if non-pacemaker?

- **Wound check documented?**

- If abnormal clinical features, has wound checked?

- **Presenting ECG rhythm and rate, and**

- If inappropriate device function seen, has ECG recorded?

- **Underlying rhythm recorded?**

- If AV conduction seen, is algorithm programmed on to minimise (or justify change in impedance, have further investigations been performed (arm movements, chest X-ray)?
- If previous AV node ablation, has CHB been confirmed?
- Programmed mode appropriate for current pacing indication?

- **Battery status recorded (Longevity / voltage / magnet rate / impedance)?**

- If discrepancy between battery parameters, has further advice been sought from manufacturer?
- At 6/12 remaining has assessment of need for device, or a referral to the CRM nurses been made?

- Changes made to maximise longevity where box change is less desirable?

- **Threshold measurements recorded?**

- Measurements verified?
- Pulse width / unipolar threshold performed if high?

- Appropriate output programmed / automatic capture management algorithm on?

- **Threshold measurements recorded?**

- Appropriate output programmed?

- **Threshold measurements recorded?**

- Appropriate output programmed?

- **% Atrial and ventricular pacing recorded?**

- **Heart rate histograms assessed and documented?**

- If inadequate heart rate variation, have settings been adjusted (or justification given if not)?
- If high ventricular rates in AF seen, has letter been sent to GP / Consultant?
- If increased proportion at max tracking rate, have settings been adjusted?

If inadequate heart rate variation, have settings been adjusted (or justification of why not)?

QA Standards

- **Appropriate lower rate limit programmed?**

- Is AMS base rate appropriate?

- **Appropriate upper track programmed?**

- **Appropriate mode programmed?**

- **Atrial**

- E

AHRE

- EGM correctly interpreted?

regulation confirmed

If MHRA advisory on device / lead, has it been adhered to?

VHRE

- If oversensing, have settings been adjusted?
- If symptomatic, has this been escalated?

- If episodes have been confirmed on ECG?

- If symptomatic SVT, has letter been sent to GP or Consultant?

- Is rate response programmed on during mode switch?

- Has the device been reprogrammed to VVIR if permanent AF (≥1 year)?

Changes documented?

Tests have been performed (arm movements / chest X-ray)?

Settings been adjusted?

Mode noted?

If symptomatic, has this been escalated?

PMTs documented?

- Is PMT prevention / termination algorithm on?

- If no algorithm available, or algorithm ineffective, has VA conduction been evaluated and PVARP adjusted?

Rate drop response / CLS programmed on for cardio-inhibitory patients?

- **If MHRA advisory on device / lead, has it been adhered to?**

- **Changes to device programming documented?**

- **Next required follow-up documented?**

- Is the time period appropriate?

- **Are reports printed?**

Data Collection

- Retrospective audit of pacemaker follow-ups

First audit cycle (100 patients)



Results disseminated



Changes implemented



Second audit cycle (100 patients)



Demographics

- 200 patients

	Audit One	Audit Two
Mean Age (\pmSD)	75 (\pm 13)	77 (\pm 10)
Mean Device Age (\pmSD)	3.5 (\pm 3)	3.6 (\pm 3)
% Male	59	56
% Dual chamber	72	69

- p = NS between groups
- Abbott (69), Boston (64), Medtronic (55), Sorin (10), Vitatron (2)

Pacing Indications


	Audit One	Audit Two
CHB	30	37
SND	26	25
AF + Brady	10	13
AF + AVNA	9	10
2:1 AVB	4	7
Mobitz I	6	4
Mobitz II	7	1
Trifascicular Block	5	2
Bifascicular Block	1	1
First Degree AVB	1	0
Cardioinhibitory VVS	1	0
Total	100	100

Changes Implemented


- **Areas for improvement**
 - Documenting heart rate histograms (36% compliance) PMTs (8%) & ULR (78%)
 - MHRA advisories (81%)
 - HF symptoms (67%)
 - Symptomatic SVT (50%)
- **Changes implemented following result feedback**
 - Follow-up form adapted
 - Advisories listed
 - HF referral letter
 - Symptomatic SVT protocol
 - Anticoagulation status for permanent AF / VVIR devices

Results


Overall QA Compliance




QA Compliance by Subgroup - Tests




QA Compliance by Subgroup - Histograms & ULR



QA Compliance by Subgroup - Episodes



QA Compliance by Subgroup



Conclusions

- National guidance for follow-up is lacking
- Local implementation has improved quality of care
- Highlighted areas of ambiguity
- QA standards do not replace clinical judgement
- Ongoing process - ?extended to ICD / CRT follow-up
- Robust QA process should be implemented across all follow-up centres



Any Questions?



Email – emma.beaney@nhs.net